
Midmark M11 - v1.0.0
Total Cycles: 1139

Name: RITTER 1
Sterilizer ID: V2178198
09/02/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.1 F 0.0
2:00 117.7 F 0.0
4:00 138.3 F 0.9
6:00 179.9 F 3.8
8:00 191.9 F 8.1
10:00 210.2 F 12.7
12:00 220.4 F 5.8
14:00 232.2 F 10.3
16:00 236.2 F 8.3
18:00 242.6 F 12.4
20:00 240.2 F 10.6
22:00 253.3 F 17.1
24:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.7 F 29.0
2:00 271.8 F 28.4
3:00 272.7 F 29.1
4:00 272.1 F 28.3

Min 271.7 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 243.5 F 0.0
10:00 184.8 F 0.0
15:00 168.6 F 0.0
20:00 163.2 F 0.0
25:00 155.4 F 0.0
30:00 150.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 25:52
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

09/02/2021 10:02 AM
ACCEPTED: X
REJECTED:
Operator: U02