
Midmark M11 - v1.0.0
Total Cycles: 3957

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 11:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.6 F 0.0
2:00 180.4 F 3.3
4:00 195.9 F 7.5
6:00 216.9 F 12.4
8:00 225.6 F 6.8
10:00 233.6 F 11.9
12:00 234.7 F 8.9
14:00 240.2 F 10.3
16:00 246.8 F 13.7
18:00 258.5 F 20.0
20:00 269.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 271.7 F 28.6
2:00 272.7 F 28.9
3:00 272.3 F 29.2
4:00 271.3 F 28.7

Min 271.3 F 28.6
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 198.1 F 0.0
10:00 198.8 F 0.0
15:00 197.7 F 0.0
20:00 197.8 F 0.0
25:00 198.3 F 0.0
30:00 195.9 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 21:43
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:01:17
CYCLE COMPLETE

10/30/2023 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____