

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1568

Name: RITTER 1  
Sterilizer ID: V2178198  
01/26/2022 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.8 F 0.0  
2:00 142.1 F 2.0  
4:00 160.1 F 5.3  
6:00 183.9 F 9.3  
8:00 210.8 F 6.7  
10:00 218.2 F 7.1  
12:00 230.1 F 11.6  
14:00 231.2 F 7.6  
16:00 243.8 F 13.2  
18:00 241.9 F 11.2  
20:00 254.0 F 17.4  
22:00 265.1 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.4 F 28.8  
2:00 271.3 F 28.2  
3:00 272.4 F 28.9  
4:00 271.7 F 28.3

Min 271.2 F 28.1  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 226.0 F 0.0  
10:00 234.5 F 0.0  
15:00 228.8 F 0.0  
20:00 227.8 F 0.0  
25:00 229.9 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 23:53  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:13  
CYCLE COMPLETE

01/26/2022 03:52 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_