
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002258
8:44 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:54
STERILIZING 3:57
136.0 C 220kPa 3:57
136.1 C 222kPa 4:27
136.4 C 224kPa 4:57
136.4 C 223kPa 5:27
136.3 C 224kPa 5:57
136.1 C 222kPa 6:27
136.4 C 225kPa 6:57
136.3 C 222kPa 7:27
136.3 C 222kPa 7:27
VENTING 7:27
AIR DRYING 8:22
CYCLE COMPLETE 36:21

Digital Signature #
02DEEBF1D1B4E871
