
Midmark M11 - v1.0.0
Total Cycles: 4065

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2023 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 189.3 F 3.2
4:00 195.6 F 7.1
6:00 211.8 F 12.0
8:00 221.7 F 6.2
10:00 229.5 F 11.1
12:00 231.4 F 7.7
14:00 244.3 F 13.0
16:00 243.8 F 12.3
18:00 256.1 F 18.6
20:00 267.0 F 25.3
22:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 271.8 F 28.9
2:00 271.7 F 28.9
3:00 271.8 F 28.9
4:00 272.1 F 28.9

Min 271.6 F 28.8
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 250.5 F 0.0
10:00 236.7 F 0.0
15:00 227.6 F 0.0
20:00 225.8 F 0.0
25:00 231.2 F 0.0
30:00 228.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

11/29/2023 01:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____