

-----  
Midmark M11 - v1.0.5  
Total Cycles: 148

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2021 10:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.6 F 0.0  
2:00 186.9 F 2.8  
4:00 190.0 F 7.1  
6:00 208.3 F 11.8  
8:00 220.7 F 7.0  
10:00 232.7 F 11.9  
12:00 234.7 F 9.5  
14:00 238.6 F 10.0  
16:00 247.7 F 14.6  
18:00 259.1 F 20.9  
20:00 269.4 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.4 F 29.5  
2:00 271.5 F 29.3  
3:00 272.2 F 29.5  
4:00 271.3 F 29.1

Min 271.3 F 28.9  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 227.1 F 0.0  
10:00 237.9 F 0.0  
15:00 231.9 F 0.0  
20:00 227.3 F 0.0  
25:00 228.9 F 0.0  
30:00 182.1 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 21:01  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

06/28/2021 12:13 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01