
Midmark M11 - v1.0.0
Total Cycles: 5273

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2025 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.8 F 0.0
2:00 85.8 F 0.0
4:00 93.6 F 0.3
6:00 126.7 F 2.8
8:00 158.7 F 5.8
10:00 184.5 F 9.4
12:00 215.6 F 13.5
14:00 218.4 F 5.2
16:00 226.7 F 8.9
18:00 237.9 F 13.3
20:00 234.7 F 8.6
22:00 243.5 F 11.9
24:00 243.2 F 11.6
26:00 255.3 F 17.9
28:00 266.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.1 F 28.3
2:00 272.1 F 28.6
3:00 271.6 F 27.9
4:00 272.6 F 28.7

Min 271.4 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 187.9 F 0.0
10:00 193.2 F 0.0
15:00 195.9 F 0.0
20:00 204.7 F 0.0
25:00 203.0 F 0.0
30:00 196.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 29:34
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:07:45
CYCLE COMPLETE

01/24/2025 10:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____