
Midmark M11 - v1.0.0
Total Cycles: 4735

Name: RITTER 1
Sterilizer ID: V2178198
07/16/2024 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.8 F 0.0
2:00 125.7 F 1.6
4:00 136.8 F 4.4
6:00 162.2 F 7.5
8:00 180.8 F 11.0
10:00 210.6 F 4.9
12:00 216.2 F 7.6
14:00 227.5 F 11.6
16:00 227.7 F 7.0
18:00 239.0 F 12.2
20:00 238.0 F 10.0
22:00 250.0 F 15.7
24:00 260.4 F 21.6
26:00 270.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.7
1:00 272.4 F 29.7
2:00 272.2 F 29.3
3:00 272.2 F 29.6
4:00 272.3 F 29.4

Min 271.0 F 28.6
Max 273.2 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 219.1 F 0.0
10:00 209.9 F 0.0
15:00 206.2 F 0.0
20:00 228.0 F 0.0
25:00 237.3 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 26:54
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:05:45
CYCLE COMPLETE

07/16/2024 02:17 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____