
Midmark M11 - v1.0.5
Total Cycles: 1190

Name: RITTER2
Sterilizer ID: V2313228
06/15/2022 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.3 F 0.0
2:00 176.3 F 3.5
4:00 185.0 F 7.7
6:00 213.7 F 13.1
8:00 229.3 F 8.4
10:00 235.9 F 11.4
12:00 240.1 F 11.9
14:00 240.9 F 11.0
16:00 254.4 F 17.9
18:00 266.4 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 271.6 F 28.8
2:00 272.3 F 29.1
3:00 272.5 F 29.4
4:00 271.8 F 29.4

Min 271.3 F 28.8
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 221.0 F 0.0
10:00 236.1 F 0.0
15:00 229.2 F 0.0
20:00 225.3 F 0.0
25:00 223.7 F 0.0
30:00 221.0 F 0.0
Drying Complete

FILLING: 1:35
HEATING: 19:30
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

06/15/2022 11:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____