
Midmark M11 - v1.0.0
Total Cycles: 3737

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 03:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 165.5 F 0.3
4:00 192.6 F 3.5
6:00 208.4 F 7.1
8:00 221.4 F 11.3
10:00 225.7 F 6.4
12:00 237.5 F 11.4
14:00 235.4 F 8.9
16:00 240.5 F 10.5
18:00 247.4 F 13.9
20:00 259.1 F 20.3
22:00 269.5 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.8
2:00 271.5 F 28.5
3:00 272.2 F 28.7
4:00 271.3 F 28.3

Min 271.3 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 207.9 F 0.0
10:00 214.9 F 0.0
15:00 212.4 F 0.0
20:00 209.2 F 0.0
25:00 207.4 F 0.0
30:00 205.9 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:00
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:01:56
CYCLE COMPLETE

09/07/2023 04:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____