
Midmark M11 - v1.0.0
Total Cycles: 2862

Name: RITTER 1
Sterilizer ID: V2178198
02/06/2023 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 187.5 F 4.2
4:00 193.1 F 8.9
6:00 211.3 F 8.8
8:00 225.4 F 8.4
10:00 236.3 F 11.6
12:00 239.4 F 11.3
14:00 238.3 F 9.7
16:00 252.4 F 16.7
18:00 264.1 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.5 F 29.3
2:00 271.9 F 28.7
3:00 271.3 F 28.6
4:00 272.4 F 29.4

Min 271.1 F 28.3
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.4 F 0.0
5:00 221.1 F 0.0
10:00 232.3 F 0.0
15:00 234.3 F 0.0
20:00 231.8 F 0.0
25:00 228.5 F 0.0
30:00 228.8 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:59
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

02/06/2023 03:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____