
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.2uS / 1.4ppm
CYCLE NUMBER 003677
8:34 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:32
STERILIZING 3:29
136.4 C 222kPa 3:29
136.4 C 222kPa 3:59
136.0 C 222kPa 4:29
136.3 C 221kPa 4:59
136.4 C 221kPa 5:29
136.3 C 216kPa 5:59
136.4 C 223kPa 6:29
136.3 C 223kPa 6:59
136.3 C 223kPa 6:59
VENTING 6:59
AIR DRYING 7:47
CYCLE COMPLETE 12:00

Digital Signature #
BC41D2F3B965DFBF
