
Midmark M11 - v1.0.0
Total Cycles: 1149

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2021 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.5 F 0.0
2:00 204.5 F 3.9
4:00 217.9 F 10.0
6:00 224.2 F 5.9
8:00 237.0 F 11.6
10:00 236.5 F 9.4
12:00 238.3 F 9.2
14:00 250.0 F 15.3
16:00 262.4 F 22.4
18:00 272.4 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.3 F 28.7
2:00 272.7 F 29.2
3:00 271.8 F 28.3
4:00 272.8 F 29.1

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 215.8 F 0.0
10:00 210.6 F 0.0
15:00 206.4 F 0.0
20:00 200.9 F 0.0
25:00 196.5 F 0.0
30:00 187.1 F 0.0
Drying Complete

FILLING: 2:26
HEATING: 18:13
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

09/08/2021 12:09 PM
ACCEPTED: X
REJECTED:
Operator: U01