

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001010  
13:41 16/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:08  
PRESSURIZING 2:12  
STERILIZING 3:15  
136.5 C 231kPa 3:15  
136.0 C 229kPa 3:45  
136.2 C 231kPa 4:15  
136.4 C 233kPa 4:45  
136.2 C 231kPa 5:15  
136.0 C 232kPa 5:45  
136.4 C 233kPa 6:15  
136.4 C 231kPa 6:45  
136.4 C 231kPa 6:45  
VENTING 6:45  
AIR DRYING 7:41  
CYCLE COMPLETE 10:15  
-----

Digital Signature #  
C332F05D9E9405C7  
-----