
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001592
12:49 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:52
STERILIZING 3:57
136.4 C 224kPa 3:57
136.2 C 223kPa 4:27
136.4 C 224kPa 4:57
136.2 C 224kPa 5:27
136.2 C 222kPa 5:57
136.1 C 221kPa 6:27
136.3 C 224kPa 6:57
136.0 C 221kPa 7:27
136.0 C 221kPa 7:27
VENTING 7:27
AIR DRYING 8:17
CYCLE COMPLETE 8:25

Digital Signature #
5948476CCD40C3BD
