

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3720

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2024 03:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.9 F 0.0  
2:00 161.9 F 2.9  
4:00 176.3 F 7.1  
6:00 199.4 F 12.1  
8:00 216.7 F 6.6  
10:00 232.3 F 12.2  
12:00 234.3 F 10.0  
14:00 235.6 F 8.7  
16:00 249.2 F 15.7  
18:00 261.5 F 22.7  
20:00 272.6 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.8 F 30.0  
2:00 271.3 F 29.3  
3:00 272.4 F 30.2  
4:00 271.9 F 29.6  
  
Min 271.2 F 29.3  
Max 273.0 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.9 F 0.0  
5:00 210.3 F 0.0  
10:00 209.5 F 0.0  
15:00 200.2 F 0.0  
20:00 192.6 F 0.0  
25:00 188.7 F 0.0  
30:00 186.7 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 20:27  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:22  
CYCLE COMPLETE

06/19/2024 04:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_