

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000920  
11:34 10/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:52  
STERILIZING 4:17  
136.4 C 231kPa 4:17  
136.4 C 233kPa 4:47  
136.4 C 232kPa 5:17  
136.1 C 232kPa 5:47  
136.4 C 232kPa 6:17  
136.2 C 230kPa 6:47  
136.3 C 232kPa 7:17  
136.4 C 233kPa 7:47  
136.4 C 233kPa 7:47  
VENTING 7:47  
AIR DRYING 8:41  
CYCLE COMPLETE 13:22

-----  
Digital Signature #  
5BDA4C9913C89CCB  
-----