
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002582
9:09 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:45
STERILIZING 3:49
136.4 C 225kPa 3:49
136.1 C 223kPa 4:19
136.3 C 224kPa 4:49
136.4 C 226kPa 5:19
136.3 C 224kPa 5:49
136.3 C 225kPa 6:19
136.4 C 225kPa 6:49
136.3 C 224kPa 7:19
136.3 C 224kPa 7:19
VENTING 7:19
AIR DRYING 8:09
CYCLE COMPLETE 14:16

Digital Signature #
1C25D30200379526
