
Midmark M11 - v1.0.5
Total Cycles: 4459

Name: RITTER2
Sterilizer ID: V2313228
03/26/2025 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.9 F 0.0
2:00 168.3 F 2.0
4:00 171.0 F 5.7
6:00 180.2 F 10.3
8:00 216.3 F 5.9
10:00 224.7 F 9.3
12:00 232.7 F 10.6
14:00 238.8 F 11.8
16:00 238.8 F 10.4
18:00 250.4 F 16.8
20:00 262.5 F 23.3
22:00 272.4 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.4
1:00 272.9 F 30.9
2:00 271.9 F 30.3
3:00 272.3 F 31.1
4:00 272.5 F 30.6

Min 271.6 F 30.2
Max 273.0 F 31.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 202.2 F 0.0
10:00 214.9 F 0.0
15:00 210.2 F 0.0
20:00 207.6 F 0.0
25:00 206.5 F 0.0
30:00 205.3 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

03/26/2025 12:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____