
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002388
14:14 16/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:51
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:51

Digital Signature #
4173A0F4DE90CDB3
