

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3798

Name: RITTER2  
Sterilizer ID: V2313228  
07/18/2024 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.6 F 0.0  
2:00 104.7 F 0.0  
4:00 113.1 F 0.1  
6:00 155.7 F 2.9  
8:00 173.8 F 6.9  
10:00 194.0 F 11.1  
12:00 216.5 F 4.9  
14:00 227.5 F 9.8  
16:00 232.7 F 9.4  
18:00 239.4 F 11.8  
20:00 238.0 F 10.1  
22:00 250.4 F 16.3  
24:00 261.6 F 22.6  
26:00 271.8 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.3  
1:00 272.7 F 30.1  
2:00 272.0 F 29.4  
3:00 271.3 F 29.5  
4:00 272.8 F 30.0

Min 271.2 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
5:00 220.2 F 0.0  
10:00 223.7 F 0.0  
15:00 210.2 F 0.0  
20:00 209.9 F 0.0  
25:00 214.7 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 26:39  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:05:04  
CYCLE COMPLETE

07/18/2024 02:36 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_