
Midmark M11 - v1.0.5
Total Cycles: 2227

Name: RITTER2
Sterilizer ID: V2313228
04/11/2023 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.9 F 0.0
2:00 80.3 F 0.0
4:00 88.5 F 0.0
6:00 112.0 F 0.4
8:00 157.0 F 3.3
10:00 171.2 F 6.9
12:00 199.4 F 10.8
14:00 215.1 F 5.8
16:00 220.7 F 8.9
18:00 233.7 F 13.0
20:00 234.3 F 9.6
22:00 237.3 F 9.6
24:00 245.0 F 13.3
26:00 256.4 F 19.5
28:00 267.0 F 26.0
30:00 273.0 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.8
1:00 272.9 F 30.3
2:00 271.9 F 29.7
3:00 272.9 F 30.4
4:00 272.2 F 29.7

Min 271.8 F 29.6
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 211.7 F 0.0
10:00 220.9 F 0.0
15:00 223.6 F 0.0
20:00 224.2 F 0.0
25:00 226.4 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:20
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

04/11/2023 08:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____