

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001703  
11:02 08/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:05  
STERILIZING 4:06  
136.4 C 222kPa 4:06  
136.1 C 222kPa 4:36  
136.4 C 224kPa 5:06  
136.4 C 225kPa 5:36  
136.4 C 225kPa 6:06  
136.2 C 219kPa 6:36  
136.4 C 225kPa 7:06  
136.4 C 220kPa 7:36  
136.4 C 220kPa 7:36  
VENTING 7:36  
AIR DRYING 8:32  
CYCLE COMPLETE 9:13

-----  
Digital Signature #  
958EBEF6EDD31ADD  
-----