
Midmark M11 - v1.0.0
Total Cycles: 4023

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.6 F 0.0
2:00 157.3 F 2.5
4:00 161.7 F 5.7
6:00 194.2 F 9.8
8:00 214.1 F 6.6
10:00 216.6 F 7.2
12:00 228.4 F 11.6
14:00 229.9 F 7.6
16:00 242.3 F 13.0
18:00 241.2 F 11.3
20:00 253.4 F 17.3
22:00 264.0 F 23.5
24:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 271.9 F 29.0
2:00 272.6 F 29.0
3:00 271.6 F 28.5
4:00 271.8 F 28.9

Min 271.3 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 220.6 F 0.0
10:00 232.7 F 0.0
15:00 227.9 F 0.0
20:00 227.0 F 0.0
25:00 226.4 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:01
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

11/15/2023 02:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____