

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 002545  
10:26 27/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:42  
PRESSURIZING 1:48  
STERILIZING 2:48  
136.0 C 220kPa 2:48  
136.0 C 221kPa 3:18  
136.3 C 220kPa 3:48  
136.5 C 224kPa 4:18  
136.1 C 223kPa 4:48  
136.4 C 221kPa 5:18  
136.4 C 224kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 224kPa 6:18  
VENTING 6:18  
AIR DRYING 7:13  
CYCLE COMPLETE 7:30  
-----

Digital Signature #  
4C5BA819209E86DA  
-----