
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001146
11:02 28/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:55
STERILIZING 5:05
136.0 C 220kPa 5:05
136.1 C 223kPa 5:35
136.3 C 225kPa 6:05
136.4 C 225kPa 6:35
136.4 C 224kPa 7:05
136.1 C 222kPa 7:35
136.3 C 223kPa 8:05
136.3 C 224kPa 8:35
136.3 C 224kPa 8:35
VENTING 8:35
AIR DRYING 9:26
CASSETTE REMOVED
CYCLE COMPLETE 10:50

Digital Signature #
FA5419B2BB61DE8A
