
Midmark M11 - v1.0.0
Total Cycles: 2902

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 176.8 F 3.8
4:00 187.6 F 8.2
6:00 207.3 F 13.3
8:00 221.6 F 7.6
10:00 235.1 F 12.7
12:00 236.3 F 9.7
14:00 237.2 F 8.9
16:00 248.4 F 14.4
18:00 260.6 F 21.2
20:00 271.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.4
3:00 272.5 F 28.9
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 229.5 F 0.0
10:00 231.5 F 0.0
15:00 228.0 F 0.0
20:00 229.8 F 0.0
25:00 229.6 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 20:36
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

02/14/2023 06:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____