

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3468

Name: RITTER2  
Sterilizer ID: V2313228  
03/21/2024 04:34 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.3 F 0.0  
2:00 173.3 F 3.2  
4:00 188.7 F 7.7  
6:00 208.8 F 12.9  
8:00 223.6 F 7.5  
10:00 235.1 F 12.9  
12:00 238.4 F 10.8  
14:00 237.7 F 9.9  
16:00 252.4 F 16.9  
18:00 264.6 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.5 F 29.4  
2:00 271.3 F 28.7  
3:00 272.4 F 29.6  
4:00 271.9 F 29.0  
  
Min 271.2 F 28.7  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 211.2 F 0.0  
10:00 205.3 F 0.0  
15:00 203.8 F 0.0  
20:00 202.8 F 0.0  
25:00 200.2 F 0.0  
30:00 197.7 F 0.0  
  
Drying Complete

FILLING: 1:34  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 00:58:57  
CYCLE COMPLETE

03/21/2024 05:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_