

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002843  
9:18 15/12/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:03  
STERILIZING 4:01  
136.4 C 222kPa 4:01  
136.3 C 223kPa 4:31  
136.4 C 222kPa 5:01  
136.3 C 220kPa 5:31  
136.2 C 221kPa 6:01  
136.1 C 221kPa 6:31  
136.4 C 220kPa 7:01  
136.2 C 219kPa 7:31  
136.2 C 219kPa 7:31  
VENTING 7:31  
AIR DRYING 8:23  
CYCLE COMPLETE 8:25

-----  
Digital Signature #  
CD4C41D6A19C2CF5  
-----