
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001875
8:01 22/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 4:03
136.5 C 223kPa 4:03
136.1 C 222kPa 4:33
136.4 C 225kPa 5:03
136.3 C 224kPa 5:33
136.2 C 222kPa 6:03
136.0 C 220kPa 6:33
136.2 C 222kPa 7:03
136.2 C 222kPa 7:33
136.2 C 222kPa 7:33
VENTING 7:33
AIR DRYING 8:23
CYCLE COMPLETE 10:35

Digital Signature #
009C5E3E46DDCAC9
