

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5391

Name: RITTER 1  
Sterilizer ID: V2178198  
03/05/2025 04:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.1 F 0.0  
2:00 146.1 F 2.7  
4:00 159.7 F 6.2  
6:00 193.9 F 10.2  
8:00 219.8 F 4.7  
10:00 215.0 F 8.1  
12:00 227.6 F 12.4  
14:00 234.0 F 8.8  
16:00 241.6 F 11.2  
18:00 243.5 F 12.3  
20:00 255.0 F 18.3  
22:00 265.5 F 24.6  
24:00 270.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 271.8 F 29.5  
2:00 271.8 F 29.5  
3:00 271.7 F 29.5  
4:00 271.8 F 29.3

Min 271.7 F 29.3  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 196.8 F 0.0  
10:00 206.3 F 0.0  
15:00 207.7 F 0.0  
20:00 204.6 F 0.0  
25:00 208.5 F 0.0  
30:00 209.2 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 24:35  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:33  
CYCLE COMPLETE

03/05/2025 05:44 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_