
Midmark M11 - v1.0.0
Total Cycles: 3115

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.0 F 0.0
2:00 149.3 F 0.0
4:00 169.2 F 2.9
6:00 197.9 F 7.3
8:00 216.2 F 12.4
10:00 225.0 F 6.8
12:00 236.0 F 11.7
14:00 234.0 F 8.5
16:00 241.6 F 10.9
18:00 245.4 F 13.0
20:00 257.3 F 19.5
22:00 268.3 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.0
2:00 271.3 F 28.6
3:00 272.4 F 29.0
4:00 271.3 F 28.3

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 204.9 F 0.0
10:00 236.8 F 0.0
15:00 231.5 F 0.0
20:00 224.1 F 0.0
25:00 227.8 F 0.0
30:00 217.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 23:15
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:25
CYCLE COMPLETE

04/14/2023 02:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____