
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001428
10:45 14/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:29
136.4 C 224kPa 4:29
136.3 C 224kPa 4:59
136.4 C 225kPa 5:29
136.1 C 222kPa 5:59
136.3 C 225kPa 6:29
136.4 C 224kPa 6:59
136.2 C 224kPa 7:29
136.4 C 225kPa 7:59
136.4 C 225kPa 7:59
VENTING 7:59
AIR DRYING 8:53
CYCLE COMPLETE 10:45

Digital Signature #
2FB2348519D06B86
