
Midmark M11 - v1.0.5
Total Cycles: 3031

Name: RITTER2
Sterilizer ID: V2313228
10/19/2023 11:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.8 F 0.0
2:00 180.9 F 3.2
4:00 197.5 F 7.9
6:00 213.8 F 12.6
8:00 219.9 F 7.5
10:00 233.7 F 12.9
12:00 237.7 F 10.8
14:00 237.7 F 9.8
16:00 251.5 F 16.6
18:00 263.1 F 23.4
20:00 272.4 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.8
2:00 271.8 F 29.2
3:00 271.2 F 29.1
4:00 272.3 F 29.8

Min 271.0 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.2 F 0.0
5:00 237.3 F 0.0
10:00 222.6 F 0.0
15:00 225.9 F 0.0
20:00 232.2 F 0.0
25:00 229.9 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

10/19/2023 12:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____