
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001194
9:56 28/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:53
PRESSURIZING 3:57
STERILIZING 5:26
136.0 C 220kPa 5:26
136.3 C 224kPa 5:57
136.0 C 221kPa 6:27
136.3 C 224kPa 6:57
136.4 C 224kPa 7:27
136.4 C 225kPa 7:57
136.3 C 223kPa 8:27
136.4 C 223kPa 8:57
136.4 C 223kPa 8:57
VENTING 8:57
AIR DRYING 9:54
CYCLE COMPLETE 12:46

Digital Signature #
E5CFBA1C50DBC0E8
