

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4141

Name: RITTER 1  
Sterilizer ID: V2178198  
12/20/2023 02:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.5 F 0.0  
2:00 180.4 F 3.6  
4:00 212.7 F 8.5  
6:00 215.2 F 8.3  
8:00 229.7 F 8.5  
10:00 240.3 F 11.9  
12:00 240.2 F 10.9  
14:00 237.9 F 9.1  
16:00 252.0 F 16.0  
18:00 263.7 F 22.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.0  
1:00 272.3 F 28.3  
2:00 272.6 F 28.7  
3:00 271.7 F 28.0  
4:00 272.6 F 28.6  
  
Min 271.5 F 27.9  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 226.0 F 0.0  
10:00 226.9 F 0.0  
15:00 226.7 F 0.0  
20:00 223.5 F 0.0  
25:00 226.5 F 0.0  
30:00 225.9 F 0.0  
  
Drying Complete

FILLING: 1:56  
HEATING: 19:59  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

12/20/2023 03:10 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_