
Midmark M11 - v1.0.5
Total Cycles: 3815

Name: RITTER2
Sterilizer ID: V2313228
07/26/2024 08:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 73.3 F 0.0
4:00 108.1 F 0.0
6:00 105.6 F 0.1
8:00 137.6 F 1.9
10:00 165.3 F 4.6
12:00 187.2 F 7.8
14:00 197.6 F 11.9
16:00 210.7 F 5.2
18:00 223.2 F 9.1
20:00 234.2 F 13.4
22:00 234.9 F 9.7
24:00 238.4 F 9.9
26:00 245.5 F 13.4
28:00 256.9 F 19.5
30:00 267.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.5 F 29.5
2:00 271.2 F 28.8
3:00 272.5 F 29.6
4:00 271.7 F 28.9

Min 271.2 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 210.9 F 0.0
10:00 202.7 F 0.0
15:00 197.9 F 0.0
20:00 195.7 F 0.0
25:00 204.0 F 0.0
30:00 210.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:28
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

07/26/2024 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____