
Midmark M11 - v1.0.0
Total Cycles: 4830

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2024 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.5 F 0.0
2:00 140.6 F 0.5
4:00 156.4 F 1.1
6:00 183.4 F 3.2
8:00 193.9 F 6.8
10:00 212.3 F 10.6
12:00 225.3 F 6.3
14:00 229.3 F 8.0
16:00 239.6 F 12.5
18:00 237.4 F 9.5
20:00 239.7 F 9.9
22:00 247.0 F 13.6
24:00 257.8 F 19.5
26:00 267.5 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 272.4 F 29.0
2:00 272.6 F 29.0
3:00 271.9 F 28.5
4:00 271.5 F 28.5

Min 271.4 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.1 F 0.0
5:00 221.8 F 0.0
10:00 227.2 F 0.0
15:00 215.9 F 0.0
20:00 222.9 F 0.0
25:00 215.1 F 0.0
30:00 169.9 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 27:37
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

08/19/2024 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____