
Midmark M11 - v1.0.0
Total Cycles: 1614

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.2 F 0.0
2:00 139.9 F 1.3
4:00 149.2 F 4.0
6:00 173.9 F 7.2
8:00 197.7 F 11.4
10:00 213.9 F 3.9
12:00 228.5 F 8.3
14:00 239.3 F 12.9
16:00 235.8 F 9.4
18:00 239.2 F 10.0
20:00 245.5 F 13.1
22:00 257.0 F 19.2
24:00 267.5 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.0 F 29.0
2:00 272.5 F 28.9
3:00 271.2 F 28.3
4:00 272.3 F 29.2

Min 271.2 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 234.7 F 0.0
10:00 212.6 F 0.0
15:00 209.7 F 0.0
20:00 200.3 F 0.0
25:00 191.1 F 0.0
30:00 194.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

02/10/2022 10:44 AM
ACCEPTED: X
REJECTED:
Operator: U02