

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005387  
8:46 20/04/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:44  
PRESSURIZING 2:47  
STERILIZING 3:41  
136.4 C 224kPa 3:41  
136.0 C 221kPa 4:11  
136.2 C 222kPa 4:41  
136.4 C 224kPa 5:11  
136.0 C 221kPa 5:41  
136.4 C 224kPa 6:11  
135.8 C 219kPa 6:41  
136.4 C 224kPa 7:11  
136.4 C 224kPa 7:11  
VENTING 7:11  
AIR DRYING 7:56  
CYCLE COMPLETE 8:04

-----  
Digital Signature #  
264CAD68EC28A4C7  
-----