

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1934

Name: RITTER2  
Sterilizer ID: V2313228  
01/13/2023 01:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.9 F 0.0  
2:00 173.1 F 2.2  
4:00 176.0 F 6.2  
6:00 194.7 F 10.9  
8:00 211.4 F 4.2  
10:00 225.8 F 9.2  
12:00 235.9 F 10.9  
14:00 239.7 F 11.6  
16:00 238.5 F 10.2  
18:00 252.0 F 16.8  
20:00 263.6 F 23.7  
22:00 271.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.7 F 29.7  
2:00 271.8 F 28.9  
3:00 272.3 F 29.8  
4:00 272.4 F 29.5

Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
5:00 221.1 F 0.0  
10:00 227.9 F 0.0  
15:00 226.9 F 0.0  
20:00 222.0 F 0.0  
25:00 217.1 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:07  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:52  
CYCLE COMPLETE

01/13/2023 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_