
Midmark M11 - v1.0.0
Total Cycles: 3927

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0

C103: HEATUP MODE

** ITEMS NOT STERILE **

10/20/2023 07:55 AM