
Midmark M11 - v1.0.0
Total Cycles: 925

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2021 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.8 F 0.0
2:00 143.3 F 1.3
4:00 155.4 F 4.2
6:00 184.5 F 8.3
8:00 218.2 F 13.1
10:00 222.6 F 8.4
12:00 227.8 F 11.4
14:00 237.3 F 10.8
16:00 234.8 F 8.3
18:00 248.5 F 14.8
20:00 259.8 F 21.2
22:00 270.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.1
1:00 272.5 F 29.5
2:00 271.5 F 29.5
3:00 271.9 F 29.3
4:00 271.9 F 29.7

Min 271.0 F 29.0
Max 273.3 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.8 F 0.0
5:00 232.6 F 0.0
10:00 235.3 F 0.0
15:00 229.4 F 0.0
20:00 227.5 F 0.0
25:00 230.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:01:14
CYCLE COMPLETE

06/22/2021 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____