
Midmark M11 - v1.0.0
Total Cycles: 4179

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2024 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.3 F 0.0
2:00 151.0 F 1.7
4:00 155.7 F 4.4
6:00 198.5 F 8.1
8:00 203.6 F 12.9
10:00 221.0 F 5.9
12:00 227.1 F 10.5
14:00 230.9 F 7.4
16:00 242.8 F 11.9
18:00 239.1 F 9.8
20:00 251.1 F 16.0
22:00 262.2 F 22.2
24:00 272.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 271.7 F 28.5
2:00 272.4 F 28.7
3:00 272.5 F 28.9
4:00 272.1 F 29.1

Min 271.3 F 28.5
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.7 F 0.0
5:00 201.3 F 0.0
10:00 189.8 F 0.0
15:00 193.6 F 0.0
20:00 202.0 F 0.0
25:00 190.5 F 0.0
30:00 188.6 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 25:07
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:03:46
CYCLE COMPLETE

01/04/2024 11:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____