
Midmark M11 - v1.0.0
Total Cycles: 5228

Name: RITTER 1
Sterilizer ID: V2178198
01/08/2025 05:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.3 F 0.0
2:00 89.3 F 0.0
4:00 109.1 F 0.0
6:00 147.1 F 0.3
8:00 176.3 F 2.5
10:00 199.0 F 6.1
12:00 211.1 F 9.5
14:00 223.0 F 12.8
16:00 227.3 F 7.4
18:00 237.8 F 11.9
20:00 235.0 F 8.5
22:00 242.1 F 11.1
24:00 244.9 F 12.5
26:00 256.8 F 18.9
28:00 267.6 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.7 F 28.0
2:00 272.0 F 28.2
3:00 272.4 F 28.5
4:00 272.7 F 28.7

Min 271.6 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 199.6 F 0.0
10:00 192.1 F 0.0
15:00 194.6 F 0.0
20:00 200.5 F 0.0
25:00 198.4 F 0.0
30:00 195.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:18
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:07:33
CYCLE COMPLETE

01/08/2025 06:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____