

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3238

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2023 08:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
2:00 190.0 F 0.0  
4:00 194.3 F 2.8  
6:00 210.7 F 8.2  
8:00 226.3 F 13.6  
10:00 232.4 F 9.3  
12:00 236.5 F 9.4  
14:00 244.1 F 13.0  
16:00 243.0 F 11.9  
18:00 255.9 F 18.6  
20:00 267.3 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 272.4 F 29.2  
2:00 272.0 F 28.6  
3:00 271.8 F 28.9  
4:00 272.6 F 29.1  
  
Min 271.5 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 220.9 F 0.0  
10:00 221.8 F 0.0  
15:00 219.4 F 0.0  
20:00 218.7 F 0.0  
25:00 217.0 F 0.0  
30:00 215.3 F 0.0  
Drying Complete

FILLING: 1:19  
HEATING: 21:17  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:52  
CYCLE COMPLETE

05/16/2023 09:25 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_