
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001330
8:59 01/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:46
STERILIZING 5:02
136.0 C 220kPa 5:02
136.2 C 221kPa 5:32
136.4 C 224kPa 6:02
136.4 C 225kPa 6:32
136.3 C 223kPa 7:02
136.3 C 224kPa 7:32
136.3 C 224kPa 8:02
136.4 C 224kPa 8:32
136.4 C 224kPa 8:32
VENTING 8:32
AIR DRYING 9:28
CYCLE COMPLETE 24:28

Digital Signature #
DA59D7B2CCD76F8A
