
Midmark M11 - v1.0.0
Total Cycles: 3809

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 187.3 F 2.5
4:00 192.1 F 6.6
6:00 210.5 F 11.2
8:00 226.6 F 5.6
10:00 231.1 F 10.5
12:00 233.3 F 7.9
14:00 243.5 F 12.9
16:00 242.6 F 11.3
18:00 254.4 F 17.3
20:00 264.9 F 23.6
22:00 270.8 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 271.9 F 28.5
2:00 272.6 F 28.8
3:00 272.4 F 29.0
4:00 271.4 F 28.5

Min 271.3 F 28.4
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 234.5 F 0.0
10:00 194.4 F 0.0
15:00 185.5 F 0.0
20:00 182.0 F 0.0
25:00 178.2 F 0.0
30:00 196.6 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:01:44
CYCLE COMPLETE

09/25/2023 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____