
Midmark M11 - v1.0.0
Total Cycles: 701

Name: MIDMARK
Sterilizer ID: V2178198
02/25/2021 07:50 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 76.2 F 0.0
4:00 81.0 F 0.0
6:00 117.0 F 0.1
8:00 150.8 F 0.3
10:00 184.5 F 3.3
12:00 204.8 F 6.9
14:00 218.2 F 11.0
16:00 223.1 F 5.5
18:00 233.4 F 9.9
20:00 235.6 F 8.7
22:00 243.8 F 12.4
24:00 241.0 F 10.7
26:00 255.3 F 17.9
28:00 267.1 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.4 F 28.6
2:00 272.3 F 28.8
3:00 271.5 F 28.0

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 223.9 F 0.0
10:00 219.3 F 0.0
15:00 219.1 F 0.0
20:00 223.0 F 0.0
25:00 230.0 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:14
STERILIZING: 3:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 01:05:58
CYCLE COMPLETE

02/25/2021 08:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____