
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006299
8:33 14/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:19
PRESSURIZING 3:21
STERILIZING 4:22
136.4 C 222kPa 4:22
136.2 C 218kPa 4:52
136.3 C 222kPa 5:22
136.1 C 219kPa 5:52
136.3 C 222kPa 6:22
136.1 C 219kPa 6:52
136.3 C 223kPa 7:22
136.3 C 219kPa 7:52
136.3 C 219kPa 7:52
VENTING 7:52
AIR DRYING 8:38
CYCLE COMPLETE 8:48

Digital Signature #
F03D0DEC0801CD90
